


Spectrophotometer Test Report

Light Source Test Report



CIE Color Parameters:

Chromaticity Coordinate: $x=0.3832$ $y=0.3832$ / $u'=0.2244$ $v'=0.5048$ ($duv=2.18e-003$)

CCT: $T_c=3973K$ Prcp WaveL: $\lambda_c=578.1nm$ Purity=30.0%

Peak WaveL: $\lambda_p=450nm$ Half Width: $\Delta\lambda=27.8nm$ Ratio: R=19.7% G=76.0% B=4.2%

Average Wave: 580nm PB=4.2298 PG=4.6836 PFCLASS: OUT=196.3604

Rendering Index: Ra=93.5

R1 =94 R2 =96 R3 =97 R4 =93 R5 =92 R6 =93 R7 =95 R8 =87

R9 =68 R10=90 R11=93 R12=71 R13=94 R14=98 R15=91

Photo Parameters:

Flux: $\Phi=3341.0(lm)$ Luminous Efficacy: 92.66(lm/W) Luminous Power: P=11.32(W)

Electrical Parameters:

U=221.0V I=0.1667A P=36.06W PF=0.979

Instrument Status:

Scan Range: 380.0nm-800.0nm

Interval: 5.0nm

Ip = 3012(G=6,D=49)

REF = 19017

TMP(PMT) = 18.7degrees centigrad/Test Mode: Fast Test

Product Type: 47134401
Instrument: PMS-80 System
Temperature: 25.0deg
Test Operator: LXF

Manufacturer: JT
Test Department: JT
Humidity: 25.0%
Test Date: 2021-04-07 12:21